

Table S1. Analytical data for the calibration curve of IN assay.

| Range (%) | Concentration ($\mu\text{g mL}^{-1}$) | Mean of the areas (a.u.) | RSD (%) |
|------------------|---|---------------------------------|----------------|
| 50 | 25 | 502856 | 0.51 |
| 60 | 30 | 603460 | 1.11 |
| 70 | 35 | 696487 | 0.10 |
| 80 | 40 | 810419 | 0.47 |
| 90 | 45 | 899034 | 0.39 |
| 100 | 50 | 993659 | 0.71 |
| 110 | 55 | 1098174 | 0.09 |
| 120 | 60 | 1204940 | 0.33 |
| 130 | 65 | 1317975 | 1.74 |
| 150 | 75 | 1494289 | 0.52 |