

**Table S3.** Analytical data for accuracy determination in IN assay.

<b>Level (%)</b>	<b>Concentration taken (<math>\mu\text{g mL}^{-1}</math>)</b>	<b>Mean peak areas (a.u.)</b>	<b>Recovery (%)</b>
80	40.80	906951	98.03
80	40.80	962348	103.76
80	40.80	938609	101.23
100	51.00	1188174	101.68
100	51.00	1218531	104.19
100	51.00	1201022	102.74
120	61.20	1421353	100.80
120	61.20	1467640	103.99
120	61.20	1443445	102.32
		Mean	102.08
		SD	1.95
		SDR (%)	1.91