

Supplementary materials

Selective separation of thorium from uranyl ions using ultrasonic-assisted solidified floating organic drop microextraction

Ghada G. El-Gamal ^a, Wael I. Mortada ^{b,*}, Mohamed M. Hassanien ^c, Amr A. Ibrahim ^a,
Yasmeen G. Abou El-Reash ^a

^a Chemistry Department, Faculty of Science, Mansoura University, Mansoura, Egypt

^b Urology and Nephrology Center, Mansoura University, Mansoura, Egypt

^c Chemistry Department, Industrial Education College, Beni-Suef University, Beni-Suef, Egypt

* Corresponding author:

Dr. Wael I. Mortada,

Urology and Nephrology Center, Mansoura University, Mansoura 35516, Egypt,

E-mail: w.mortada@mans.edu.eg

Table S1 Instrumental parameters for the determination of Th by ICP-OES

RF power (W)	1200
Plasma gas flow rate (L min ⁻¹)	12
Auxiliary gas flow rate (L min ⁻¹)	1
Nebulizer gas flow rate (L min ⁻¹)	0.7
Delay time (s)	15
Integration time (s)	3
Replicates	3
Wavelength (nm)	Th 283.730
