

Electronic Supplementary Material (ESI) for Journal of Analytical Atomic Spectrometry.
This journal is © The Royal Society of Chemistry 2023

Input				Calculations					
<i>comment</i>	<i>raster</i>	<i>I-Fco</i>	<i>time</i>	<i>time</i>	<i>area</i>	<i>charge</i>	<i>fluence</i>	<i>dose</i>	<i>dose</i>
	<i>[μm]</i>	<i>[pA]</i>	<i>[s]</i>	<i>[hh:mm:ss]</i>	<i>[m²]</i>	<i>[C]</i>	<i>[C/m²]</i>	<i>[ions/m²]</i>	<i>[ions/cm²]</i>
Fluence 1C/m ²	10	10	10	00:00:10	1.0E-10	1.0E-10	1.0	6.2E+18	6.2E+14
Implantation check	10	12	500	00:08:20	1.0E-10	6.0E-09	60.0	3.7E+20	3.7E+16
Implantion, high current, 60C/m ²	50	260	580	00:09:40	2.5E-09	1.5E-07	60.3	3.8E+20	3.8E+16
Measurement, 1 plane, 256 pixels	30	2	66	00:01:06	9.0E-10	1.3E-10	0.15	9.1E+17	9.1E+13
Measurement, 20 planes, 256 pixels	30	2	1312	00:21:52	9.0E-10	2.6E-09	2.9	1.8E+19	1.8E+15
Implantation steady state	50	260	1445	00:24:05	2.5E-09	3.8E-07	150.3	9.4E+20	9.4E+16