

Table SI1 Results obtained in the analysis of the Certified Reference Material (EnviroMAT™ EU-L Waste Water, Low).

Element	Consensus value		
	(mg/L)	Found value (mg/L)	Recovery (%)
As	0.0792	0.0873	110
Ba	0.123	0.117	95
Cd	0.0229	0.0266	116
Co	0.0805	0.0705	88
Cr	0.0612	0.0557	91
Cu	0.107	0.099	93
Fe	0.0654	0.0531	81
Mn	0.121	0.118	98
Mo	0.0404	0.0366	91
Ni	0.0816	0.0795	97
Pb	0.0413	0.0403	98
Sb	0.0186	0.0232	125
Se	0.0265	0.0333	126
Tl	0.0824	0.0943	114
V	0.0484	0.0482	100
Zn	0.0292	0.0296	101