

**Supplementary Material for:**

**Direct Immunoassay on Polyester Microwell Plate for Colorimetric Detection of Spike Protein in Swab and Saliva Samples**

Nikaele S. Moreira,<sup>a</sup> Thaisa A. Baldo,<sup>a</sup> Lucas C. Duarte,<sup>a,e</sup> Leonardo Lopes-Luz,<sup>b</sup> Karoliny A. Oliveira,<sup>a</sup> Paulo F. N. Estrela,<sup>a</sup> Amanda M. Simões,<sup>a</sup> Samira Bührer-Sékula,<sup>bc</sup> Gabriela R. M. Duarte,<sup>ac</sup> and Wendell K. T. Coltro<sup>\*acd</sup>

<sup>a</sup>Instituto de Química, Universidade Federal de Goiás, 74690-900, Goiânia, GO, Brazil.

<sup>b</sup>Instituto de Patologia Tropical e Saúde Pública, Centro Multusuário de Bioinsumos e Tecnologias em Saúde, Universidade Federal de Goiás, 74605-050, Goiânia, GO, Brazil.

<sup>c</sup>Innovation Hub in Point-of-Care Technologies, 74690-900, Goiânia, GO, Brazil.

<sup>d</sup>Instituto Nacional de Ciência e Tecnologia de Bioanalítica, 13084-971, Campinas, SP, Brazil.

<sup>e</sup>Instituto Federal de Educação, Ciência e Tecnologia de Goiás – Campus Inhumas, 75402-556, Inhumas, GO, Brazil

**\*Corresponding Author:**

Professor Wendell K. T. Coltro

Instituto de Química, Universidade Federal de Goiás


Campus Samambaia, 74690-900

Goiânia, GO, Brazil


Fax: +55 62 3521 1127

E-mail: [wendell@ufg.br](mailto:wendell@ufg.br)

**ORCID:** <http://orcid.org/0000-0002-4009-2291>



**Figure S1.** (A) Evaluation of the image percentage scale in image acquisition time using the scanner. Colorimetric responses obtained from the S protein detection reaction on the microplate were analyzed in the color systems RGB (B) and CMYK (C) of different color channels.



**Figure S2.** Optimization of sensitization step of direct ELISA for detection of SARS-CoV-2. The concentration of S-protein used was  $1.5 \mu\text{g mL}^{-1}$ .

**Table S1.** Evaluation of repeatability intraplate and interplate.

[S protein] $\mu\text{g. mL}^{-1}$ (n=5)	Intraplate (1 plate)	RSD (%) Interplate (3 plates)
<b>1.5</b>	3.7	7.7
<b>2.5</b>	7.2	5.9
<b>5.0</b>	7.8	2.3
<b>10.0</b>	3.7	5.2

**Table S2:** Results of human Saliva and Swab samples tested using the gold standard RT-qPCR assay and the developed direct ELISA method on a polyester plate.

Sample	Sample type	Classification gold std.	direct ELISA					Classification
			1 <sup>st</sup>	2 <sup>nd</sup>	3 <sup>rd</sup>	Average (A.U.)	Std dev.	
1	Saliva	Negative	4.80	5.10	4.48	<b>4.79</b>	<b>0.31</b>	Negative
2	Saliva	Negative	4.23	4.06	4.35	<b>4.21</b>	<b>0.15</b>	Negative
3	Saliva	Negative	5.21	5.1	5.32	<b>5.21</b>	<b>0.11</b>	Negative
4	Saliva	Negative	3.24	3.05	3.75	<b>3.35</b>	<b>0.36</b>	Negative
5	Saliva	Negative	4.31	4.48	5.00	<b>4.60</b>	<b>0.36</b>	Negative
6	Saliva	Negative	5.20	5.33	5.34	<b>5.29</b>	<b>0.08</b>	Negative
7	Saliva	Negative	6.23	6.49	5.48	<b>6.07</b>	<b>0.52</b>	Positive
8	Swab	Negative	4.76	3.49	3.84	<b>4.03</b>	<b>0.66</b>	Negative
9	Swab	Negative	4.30	4.45	4.10	<b>4.28</b>	<b>0.18</b>	Negative
10	Swab	Negative	3.16	3.44	3.1	<b>3.23</b>	<b>0.18</b>	Negative
11	Swab	Negative	5.42	5.24	4.47	<b>5.04</b>	<b>0.50</b>	Negative
12	Swab	Negative	3.28	3.08	3.32	<b>3.23</b>	<b>0.13</b>	Negative
13	Swab	Negative	2.57	2.56	2.74	<b>2.62</b>	<b>0.10</b>	Negative
14	Swab	Negative	4.31	4.84	4.69	<b>4.61</b>	<b>0.27</b>	Negative
15	Swab	Negative	4.23	4.62	4.52	<b>4.46</b>	<b>0.20</b>	Negative
16	Swab	Negative	3.24	3.40	4.32	<b>3.65</b>	<b>0.58</b>	Negative
17	Saliva	Positive	6.23	6.43	6.34	<b>6.33</b>	<b>0.10</b>	Positive
18	Saliva	Positive	6.70	7.20	6.19	<b>6.70</b>	<b>0.51</b>	Positive
19	Saliva	Positive	6.06	6.58	6.79	<b>6.48</b>	<b>0.38</b>	Positive
20	Saliva	Positive	6.24	6.94	6.6	<b>6.59</b>	<b>0.35</b>	Positive
21	Saliva	Positive	6.37	6.96	7.36	<b>6.90</b>	<b>0.50</b>	Positive
22	Saliva	Positive	6.72	6.92	6.82	<b>6.82</b>	<b>0.10</b>	Positive
23	Saliva	Positive	10.0 <sup>c</sup>	9.37	10.76	<b>10.06</b>	<b>0.70</b>	Positive
24	Saliva	Positive	18.7 <sup>c</sup>	17.08	17.61	<b>17.80</b>	<b>0.83</b>	Positive
25	Saliva	Positive	14.7 <sup>c</sup>	13.45	14.07	<b>14.08</b>	<b>0.64</b>	Positive
26	Saliva	Positive	13.6 <sup>c</sup>	12.42	12.94	<b>13.00</b>	<b>0.61</b>	Positive
27	Saliva	Positive	9.92	8.56	9.23	<b>9.24</b>	<b>0.68</b>	Positive
28	Swab	Positive	9.84	9.12	10.21	<b>9.72</b>	<b>0.55</b>	Positive
29	Swab	Positive	7.22	7.17	7.24	<b>7.21</b>	<b>0.04</b>	Positive
30	Swab	Positive	6.22	6.36	6.10	<b>6.23</b>	<b>0.13</b>	Positive
31	Swab	Positive	22.7 <sup>c</sup>	20.96	21.93	<b>21.87</b>	<b>0.89</b>	Positive
32	Swab	Positive	9.57	8.36	9.09	<b>9.01</b>	<b>0.61</b>	Positive
33	Swab	Positive	9.81	9.13	10.94	<b>9.96</b>	<b>0.91</b>	Positive
34	Swab	Positive	7.28	6.83	6.35	<b>6.82</b>	<b>0.47</b>	Positive
35	Swab	Positive	7.01	7.97	7.34	<b>7.44</b>	<b>0.49</b>	Positive
36	Swab	Positive	4.61	4.15	4.75	<b>4.50</b>	<b>0.31</b>	Negative
37	Swab	Positive	6.28	6.51	6.35	<b>6.38</b>	<b>0.12</b>	Positive

**Table S3.** Cost estimate for microwell plate immunoassay on polyester substrates

Consumables	Quantity	Cost (US\$)	Manufactured plate quantity	Total cost/ plate (US\$)
Thermal laminating Sheets (A4)	1 unit	0.53	2.67	0.20
<b>Plate total</b>				<b>0.20</b>
Reagents	Concentration	Vol./microwell ( $\mu$ L)	Cost (US\$)/microwell	Total cost (US\$)/ plate
Blocking solution	1% (v:v)	2	$1.41 \times 10^{-4}$	0.013
Biotinylated detection antibody	2.0 ( $\mu$ g. mL <sup>-1</sup> )	2	$1.39 \times 10^{-2}$	1.33
Streptavidin-HRP	1:1000 (v:v)	2	$1.25 \times 10^{-3}$	0.12
TMB Liquid Substrate, for ELISA	-	2	$2.16 \times 10^{-3}$	0.21
Step washing	10 mmol.L <sup>-1</sup> PBS and 0.05% Tween 20	Volume/step - 10 mL	$2.67 \times 10^{-2}$	0.11
<b>Total reagents</b>				<b>1.78</b>
<b>Total cost (US\$)/ plate immunoassay</b>				<b>1.98</b>
<b>Total cost (US\$)/ microwell</b>				<b>0.021</b>