

Table S1. Contents of matrix / Ag ratios of BCR-2 and AGV-2 after four-step chromatography process

Matrix / Ag ratio	BCR-2	AGV-2
Ti:Ag	0.12	0.04
Fe:Ag	0.29	0.07
Cu:Ag	0.18	0.30
Zn:Ag	0.00	0.04
Ge:Ag	0.00	0.04
Rb:Ag	0.44	0.00
Nb:Ag	0.02	0.01
Zr:Ag	0.00	0.00
Cd:Ag	0.00	0.00

Table S2. Synthetic solutions with different matrix element ratios and Ag isotope compositions

Synthetic		Na ⁺	Mg ⁺	Al ⁺	S	K ⁺	Ca ⁺	Ti	Mn	Fe ⁺	Cu ⁺	Zn ⁺	Ga	As	Se	Sr	Y	Zr	Nb	Mo	Ag [#]	Cd	Sn	δ ¹⁰³ Ag	2SD	N
I	-	-	2	-	-	-	-	20	-	2	1	1	-	20	-	20	20	10	10	20	1	30	10	0.02	0.05	6
II	1.0	1.5	4.0	20	1.5	12	5	10	14	5.0	5.0	10	20	20	10	10	-	-	100	2	200	15	0.03	0.04	12	
III	1.0	1.5	4.0	5.0	1.5	12	2	2	14	0.5	0.1	10	10	20	10	10	-	-	50	0.5	200	15	0.03	0.04	12	

* The unit of major elements is wt.%. The unit of trace elements is µg/g. ^a The reference material SRM978a.

2SD is the 2 standard deviation for the synthetic solution ($n \geq 3$). N is the number of repeated measurements.

Table S3. Ag isotope compositions of the doping solutions analyzed by MC-ICP-MS

Molar element ratio	$\delta^{109}\text{Ag}_{\text{SRM978a}} (\text{\textperthousand})$	2SD (n=3)
Fe:Ag = 1:10	0.00	0.03
Fe:Ag = 1:1	0.07	0.05
Fe:Ag = 5:1	0.19	0.05
Cu:Ag = 1:2	0.00	0.04
Cu:Ag = 1:1	0.03	0.04
Cu:Ag = 2:1	0.17	0.02
Zn:Ag = 1:2	0.00	0.03
Zn:Ag = 1:1	0.01	0.04
Zn:Ag = 2:1	-0.26	0.04
Zr:Ag = 1:100	-0.01	0.02
Zr:Ag = 1:10	-0.08	0.05
Zr:Ag = 1:2	-0.39	0.04
Nb:Ag = 1:100	-0.02	0.01
Nb:Ag = 2:10	0.06	0.04
Nb:Ag = 1:2	0.26	0.03
Cd:Ag = 1:1	0.01	0.04
Cd:Ag = 5:1	-0.01	0.04
Cd:Ag = 10:1	0.08	0.06

Figure S1. Elution curves of the Ag purification procedure using different rock standards, BCR-2 (a) and AGV-2 (b) with the same separation step as Figure 1. Elution curves of Ag purification with AG50 W-X8 resin as eluted at 1 mol/L (c) and 2 mol/L (d) HNO₃.




Figure S2. The effect of Zr and Nb doping on $\delta^{109}\text{Ag}_{\text{SRM978}}$. $\delta^{109}\text{Ag}_{\text{SRM978}}$ show an extreme drifting even when SRM978a was doped with tiny amounts of Zr and Nb solutions.





Figure S3. The Ag isotope data of the in-house laboratory standards ICP-Ag and USTC-Ag analyzed by MC-ICP-MS over 3 years.



Error bars represent the 2SD uncertainties for each measurement of 0.05‰ ($n \geq 3$). The dashed lines represent the average $\delta^{109}\text{Ag}_{\text{SRM978a}}$ values of both.