

Supplementary Material

(RP)UHPLC/UV analytical method to quantify dsRNA during mRNA vaccine manufacturing process

Sara Sousa Rosa ^{a,b,c #}, Shuran Zhang ^{a #}, Yustika Sari ^a, Marco PC Marques ^{a *}

^a Department of Biochemical Engineering, University College London, Gordon Street, London,

WC1E 6BT, United Kingdom

^b Department of Bioengineering, iBB—Institute for Bioengineering and Biosciences, Instituto Superior Técnico, Universidade de Lisboa, Lisboa, Portugal

^c Associate Laboratory i4HB—Institute for Health and Bioeconomy, Instituto Superior Técnico, Universidade de Lisboa, Lisboa, Portugal


These authors contributed equally to this work

*Correspondence:


Dr. Marco P.C. Marques

University College London, Biochemical Engineering, Bernard Katz Building, Gower Street, London, WC1E 6BT, United Kingdom

e-mail: marco.marques@ucl.ac.uk



Supplementary Figure S1. Residual plots for the calibration curves used to calculated dsRNA concentration using the HPLC (a), the Dot Blot (b) and ELISA (c).



Supplementary Figure S2. Residual plots for the dsRNA spiking studies. Dot Blot results using 0.1 g L⁻¹ (a) and 0.4 g L⁻¹ dsRNA (b). HPLC results using 0.1 g L⁻¹ (c) and 0.4 g L⁻¹ of dsRNA (d).

Supplementary Table S1. Dot Blot Tukey test results for high dsRNA concentration (0.4 g.L⁻¹). Concentration used for DNA correspond to 10 and 90 nM and to NTPs to 2 and 7.6 mM, for low and high, respectively.

		Confidence Interval at 95%			
Factor 1	Factor 2	Mean difference	Lower bound	Upper bound	Adjusted P value
DNA_high	DNA_low	0.169	-0.063	0.400	0.228
	NTP_high	0.319	0.108	0.531	0.002
	NTP_low	0.042	-0.170	0.253	0.975
	RNA_high	0.008	-0.212	0.228	1.000
DNA_low	NTP_low	0.151	-0.061	0.362	0.246
	DNA_low	-0.127	-0.338	0.085	0.404
	RNA_high	-0.161	-0.381	0.059	0.223
NTP_low	NTP_high	-0.277	-0.467	-0.088	0.002
RNA_high	NTP_high	-0.312	-0.510	-0.113	0.001
	NTP_low	-0.034	-0.233	0.164	0.985

Supplementary Table S2. Dot BlotTukey test results for low dsRNA concentration. (0.1 g.L⁻¹). Concentration used for DNA correspond to 10 and 90 nM and to NTPs to 2 and 7.6 mM, for low and high, respectively.

		Confidence Interval at 95%			
Factor 1	Factor 2	Mean Difference	Lower bound	Upper bound	Adjusted P value
DNA_high	DNA_low	-0.004	-0.052	0.044	0.998
	NTP_high	-0.013	-0.058	0.032	0.864
	NTP_low	-0.022	-0.076	0.031	0.663
	RNA_low	-0.034	-0.082	0.014	0.213
DNA_low	NTP_low	-0.009	-0.054	0.036	0.965
	DNA_low	-0.018	-0.071	0.036	0.808
	RNA_low	-0.03	-0.078	0.018	0.322
NTP_low	NTP_high	-0.009	-0.060	0.042	0.974
RNA_low	NTP_high	-0.021	-0.066	0.024	0.571
	NTP_low	-0.012	-0.065	0.042	0.947

Supplementary Table S3. HPLC Tukey test results for high dsRNA concentration (0.4 g.L⁻¹). Concentration used for DNA correspond to 10 and 90 nM and to NTPs to 2 and 7.6 mM, for low and high, respectively.

		Confidence Interval at 95%			
Factor 1	Factor 2	Mean Difference	Lower bound	Upper bound	Adjusted P value
DNA_high	DNA_low	-0.014	-0.139	0.111	0.996
	NTP_high	-0.041	-0.157	0.075	0.775
	NTP_low	0.02	-0.105	0.145	0.984
	RNA_high	0.014	-0.129	0.156	0.998
DNA_low	NTP_low	-0.028	-0.153	0.098	0.950
	DNA_low	0.034	-0.100	0.168	0.919
	RNA_high	0.027	-0.122	0.177	0.974
NTP_low	NTP_high	0.061	-0.064	0.187	0.534
RNA_high	NTP_high	0.055	-0.087	0.197	0.724
	NTP_low	-0.007	-0.156	0.143	1.000

Supplementary Table S4. HPLC Tukey test results for low dsRNA concentration (0.1 g.L⁻¹). Concentration used for DNA correspond to 10 and 90 nM and to NTPs to 2 and 7.6 mM, for low and high, respectively.

Factor 1	Factor 2	Confidence Interval at 95%			
		Mean Difference	Lower bound	Upper bound	Adjusted P value
DNA_high	DNA_low	0.011	-0.011	0.033	0.503
	NTP_high	-0.003	-0.027	0.021	0.992
	NTP_low	0.007	-0.017	0.031	0.874
	RNA_low	0.009	-0.018	0.036	0.824
DNA_low	NTP_low	-0.014	-0.038	0.010	0.352
	DNA_low	-0.004	-0.028	0.020	0.976
	RNA_low	-0.002	-0.029	0.025	0.998
NTP_low	NTP_high	0.010	-0.015	0.036	0.710
RNA_low	NTP_high	0.012	-0.017	0.040	0.666
	NTP_low	0.002	-0.027	0.030	0.999