

Figure S1. Representative SRM chromatograms by liquid chromatography-tandem mass spectrometry (LCMS/MS) monitored in positive ion mode in Panels **1**, **2**, **3** and **4**-Blank sample with Milli-Q water, without deuterated IS and EPA 521 mixture, extracted under similar experimental conditions, **1**-NDEA (m/z 103.1· m/z 75.1), **2**-NPIP (m/z 115.1 · m/z 69.1), **3**-NDPA (m/z 131.1· m/z 89.1), **4**-NDBA (m/z 159.1 · m/z 57.1). *Inset* of panel **1** and Panels **5** and **6**- Solvent blanks for NDEA, NDPA and NDBA with similar parent to product ion transitions as in **1**, **3** and **4** respectively.




Figure S2. Representative SRM chromatograms by liquid chromatography-tandem mass spectrometry (LCMS/MS) monitored in positive ion mode in methanol solvent wash run *panels* **1**-NDMA (m/z 75.1· m/z 58.1), **2**-NDMA- d_6 (m/z 81.088· m/z 46.0), **3**-NDEA (m/z 103.1· m/z 75.1), **4**-NDEA- d_{10} (m/z 113.175 · m/z 81.1), **5**-NPYR (m/z 101.1· m/z 55.1), **6**-NMEA (m/z 89.1· m/z 61.1), **7**-NMOR- d_8 (m/z 125.1· m/z 95.0), **8**-NPIP (m/z 115.1· m/z 69.1), **9**-NPIP- d_{10} (m/z 125.2· m/z 78.0), **10**-NDPA (m/z 131.1· m/z 89.1).

Table S1. Precision for injection repeatability and %CV for (**A**) standard solution of *N*-Nitrosamines, EPA 521 mixture (4 ng) and (**B**) urine spiked with EPA 521 mix (4 ng) following extraction.

Compound	A		B	
	Amount (ng)	% CV	Amount (ng)	% CV
NDMA	3.85 ±0.17	4.49	1.72 ±0.10	6.74
NDEA	4.34 ±0.22	5.07	1.98 ±0.08	4.10
NMEA	4.53 ±0.13	2.79	1.84 ±0.13	7.07
NPYR	4.82 ±0.13	2.79	1.69 ±0.10	6.06
NPIP	4.11 ±0.09	2.27	1.93 ±0.17	8.80
NDPA	4.06 ± 0.23	5.57	1.94 ±0.13	6.80
NDBA	4.44 ± 0.17	3.93	1.92 ±0.13	6.95