

Supplementary Information

Quantification Capabilities of N₂ MICAP-MS with Solution Nebulization and Aerosol Desolvation

Monique Kuonen, Bodo Hattendorf and Detlef Günther*

*Correspondence: Detlef Günther, E-mail: guenther@inorg.chem.ethz.ch, Laboratory of Inorganic Chemistry, Department of Chemistry and Applied Biosciences, ETH Zurich, Vladimir – Prelog – Weg 1, 8093 Zurich, Switzerland

The following supplementary information supports the corresponding paper with additional figures (Figure S1 – S6) and tables (Table S1 – S8).

List of Figures

S1	Overview of contained certified elements in reference samples	SI-2
S2	Enhancement due to solvent load	SI-3
S3	Quantification of SLRS-6 with solution nebulization	SI-8
S4	Quantification of SLRS-6 with desolvation.....	SI-8
S5	Quantification of SLRS-6 with solution nebulization	SI-9
S6	Quantification of SLRS-6 with desolvation.....	SI-9

List of Tables

S1	Mass fractions of the reference samples and the multi-element stock solution	SI-2
S2	Enhancement and suppression due to solvent load	SI-3
S3	Plasma gas temperatures at fixed nebulizer gas flow rate.....	SI-4
S4	Ca matrix tolerance of the N ₂ MICAP-MS	SI-4
S5	Abundance normalized sensitivities for selected elements	SI-5
S6	Limits of Detection and Limits of Quantification	SI-5
S7	Quantification overview for NIST 1643f with twofold dilution.....	SI-6
S8	Quantification overview for NIST 1643f with tenfold dilution	SI-7




Figure S1: An overview of the certified elements contained in the two water reference materials. The colours in the periodic table indicate which elements were certified for NIST SRM 1643f (dark blue), SLRS-6 (green) or both (light blue).

Table S1: Certified reference mass fraction values for SLRS-6 and NIST 1643f are given with the mass fractions of the multi-element stock solution (MES), which was used to prepare the external calibration by diluting it 4 – 400-fold. The mass fractions in the MES were chosen so that its 20-fold dilution results in similar mass fractions to those of the SLRS-6, NIST 1643f, or the five-fold dilution thereof. For elements that had at least one order of magnitude difference in the mass fractions of the two water reference materials SLRS-6 and NIST 1643f, additional multi-element solutions were prepared that matched the reference samples. All mass fractions are reported in $\mu\text{g kg}^{-1}$.

Element	SLRS-6		NIST 1643f		MES
Li	- ^a		16.4	± 0.4	336
Be	0.0066 ^b	± 0.0022 ^b	13.53	± 0.11	272
B	- ^a		150	± 7	10000
Na	2770	± 220	18640	± 240	64500
Mg	2140	± 60	7380	± 60	35600
Al	33.9	± 2.2	132.5	± 1.2	600
K	650	± 50	1913	± 9	9848
Ca	8770	± 200	29140	± 300	144300
V	0.352	± 0.006	35.71	± 0.27	74.8
Cr	0.252	± 0.012	18.32	± 0.1	39.3
Mn	2.12	± 0.1	36.7	± 0.6	93.7
Fe	85	± 4	92.5	± 0.8	1770
Co	0.053 ^b	± 0.012 ^b	25.05	± 0.17	495
Ni	0.617	± 0.022	59.2	± 1.4	119
Cu	24	± 1.8	21.44	± 0.7	451
Zn	1.76	± 0.12	73.7	± 1.7	164
As	0.57	± 0.08	56.8	± 0.4	119
Se	- ^a		11.58	± 0.08	239
Rb	- ^a		12.51	± 0.12	245
Sr	40.7	± 0.3	311	± 18	1020
Mo	0.215	± 0.018	114.2	± 1.7	227
Ag	- ^a		0.960	± 0.008	19.0
Cd	0.0063	± 0.0014	5.83	± 0.13	118
Sb	0.338	± 0.006	54.9	± 0.4	109
Te	- ^a		0.967	± 0.008	18.9
Ba	14.3	± 0.5	513	± 7	1150
Tl	- ^a		6.82	± 0.03	139
Pb	0.17	± 0.026	18.30	± 0.08	39.0
Bi	- ^a		12.50	± 0.10	247
U	0.070	± 0.003		- ^a	1.38

^a not available; ^b not certified




Figure S2: Enhancement and suppression of the desolvated aerosol signal intensity compared to the signal intensity when the desolvator was deactivated. Data is shown for nebulizer gas flow rates of 800 mL min⁻¹, 900 mL min⁻¹ and 1000 mL min⁻¹.

Table S2: Enhancement and suppression due to solvent load in dependence on the nebulizer gas flow rate. The ratio of the signal intensities from the setup with the activated to the deactivated desolvator is given for different nebulizer gas flow rates (reported in mL min⁻¹).

Isotope	800	825	850	875	900	925	950	975	1000
⁷ Li ⁺	1.13	1.09	0.88	1.04	0.93	0.85	0.71	0.77	0.74
⁹ Be ⁺	1.91	1.87	1.76	1.81	1.89	1.90	1.68	1.89	1.84
²³ Na ⁺	1.17	1.00	1.01	1.07	1.02	0.97	0.82	0.83	0.78
²⁴ Mg ⁺	1.25	1.20	1.15	1.13	1.11	1.08	0.92	0.95	0.91
³⁹ K ⁺	1.19	1.06	1.04	1.01	1.01	0.96	0.87	0.85	0.80
⁵⁹ Co ⁺	1.31	1.21	1.26	1.15	1.15	1.18	1.14	1.14	1.12
⁶⁶ Zn ⁺	1.68	1.73	1.77	2.21	2.19	2.39	2.15	2.02	2.10
⁷² Ge ⁺	1.26	1.19	1.15	1.14	1.11	1.13	1.16	1.12	1.05
⁷⁵ As ⁺	2.06	1.98	2.07	2.15	2.20	2.15	2.36	2.31	2.29
⁸⁰ Se ⁺	2.42	2.67	2.90	2.91	2.97	3.07	3.21	3.13	3.20
⁸⁵ Rb ⁺	0.96	0.93	0.87	0.83	0.81	0.79	0.75	0.74	0.69
⁸⁸ Sr ⁺	0.96	0.94	0.89	0.85	0.81	0.82	0.78	0.78	0.74
¹¹¹ Cd ⁺	1.38	1.35	1.34	1.35	1.38	1.39	1.48	1.48	1.46
¹¹⁵ In ⁺	0.89	0.85	0.85	0.80	0.78	0.78	0.76	0.75	0.74
¹²⁸ Te ⁺	1.53	1.53	1.58	1.60	1.62	1.64	1.79	1.76	1.86
¹³³ Cs ⁺	0.84	0.80	0.79	0.76	0.73	0.73	0.70	0.67	0.66
¹³⁷ Ba ⁺	0.86	0.81	0.78	0.76	0.75	0.72	0.72	0.67	0.69
¹⁴⁰ Ce ⁺	0.84	0.83	0.80	0.78	0.74	0.74	0.71	0.72	0.72
¹⁷⁵ Lu ⁺	0.86	0.84	0.79	0.76	0.74	0.75	0.73	0.75	0.72
²⁰⁸ Pb ⁺	0.79	0.78	0.75	0.72	0.71	0.70	0.71	0.67	0.68
²³² Th ⁺	0.89	0.82	0.87	0.79	0.82	0.81	0.83	0.85	0.85
²³⁸ U ⁺	0.85	0.80	0.82	0.75	0.76	0.76	0.77	0.75	0.73
¹⁴ N ¹⁶ O ⁺	0.62	0.66	0.67	0.70	0.72	0.75	0.78	0.79	0.77

Table S3: Calculated plasma temperatures obtained from the pressure reduction method, the Longerich method, and the method from Houk and Praphairaksit (H&P). Temperatures were calculated for solution nebulization (SN) and desolvated (Des) sample introduction for both the N₂ MICAP and the Ar ICP at a nebulizer gas flow rate of 900 mL min⁻¹.

Method	N ₂ MICAP		Ar ICP	
	SN	Des	SN	Des.
Pressure reduction	5'220 K	5'400 K	5'780 K	5'850 K
Longerich	12'850 K	16'400 K	13'170 K	12'460 K
H&P LaO ⁺ / La ⁺	6'350 K	6'790 K	6'620 K	6'950 K
H&P CeO ⁺ / Ce ⁺	5'910 K	6'250 K	6'060 K	6'400 K
H&P PrO ⁺ / Pr ⁺	6'200 K	6'580 K	6'080 K	6'390 K

Table S4: Ca matrix tolerance of the N₂ MICAP-MS. The signal suppression of a solution with 100 mg kg⁻¹ compared to a solution with 1 mg kg⁻¹ is shown for solution nebulization (SN) and desolvated sample introduction (Des). Suppression values are given for raw ion signals and after normalization to ¹¹⁵In as internal standard (w. IS).

Isotope	SN	SN w. IS	Des	Des w. IS
⁷ Li ⁺	- 24 %	- 13 %	- 72 %	- 41 %
⁹ Be ⁺	- 02 %	+ 15 %	- 60 %	- 18 %
¹¹ B ⁺	- 07 %	+ 05 %	+ 165 %	+ 436 %
²⁴ Mg ⁺	- 21 %	- 11 %	- 64 %	- 27 %
²⁷ Al ⁺	- 29 %	- 19 %	- 61 %	- 20 %
⁵¹ V ⁺	- 13 %	- 02 %	- 57 %	- 12 %
⁵² Cr ⁺	- 18 %	- 07 %	- 59 %	- 16 %
⁵⁵ Mn ⁺	- 20 %	- 10 %	- 60 %	- 20 %
⁵⁶ Fe ⁺	- 14 %	+ 11 %	- 55 %	- 08 %
⁵⁹ Co ⁺	- 10 %	+ 01 %	- 57 %	- 12 %
⁶⁰ Ni ⁺	- 04 %	+ 17 %	- 53 %	- 04 %
⁶³ Cu ⁺	- 17 %	- 06 %	- 58 %	- 15 %
⁶⁶ Zn ⁺	- 10 %	+ 02 %	- 52 %	- 01 %
⁷¹ Ga ⁺	- 18 %	- 08 %	- 56 %	- 10 %
⁷⁵ As ⁺	- 12 %	- 01 %	- 51 %	+ 01 %
⁸⁰ Se ⁺	- 15 %	- 04 %	- 41 %	+ 20 %
⁸⁵ Rb ⁺	- 25 %	- 15 %	- 54 %	- 06 %
⁹⁸ Mo ⁺	- 07 %	+ 05 %	- 51 %	+ 01 %
¹⁰⁷ Ag ⁺	- 15 %	- 04 %	- 53 %	- 02 %
¹¹¹ Cd ⁺	- 04 %	+ 09 %	- 51 %	+ 02 %
¹²⁸ Te ⁺	- 02 %	+ 12 %	- 46 %	+ 12 %
¹³⁷ Ba ⁺	- 16 %	- 05 %	- 51 %	+ 01 %
²⁰⁸ Pb ⁺	- 16 %	- 06 %	- 54 %	- 07 %
²⁰⁹ Bi ⁺	- 10 %	+ 01 %	- 53 %	- 04 %
²³⁸ U ⁺	- 06 %	+ 06 %	- 54 %	- 05 %

Table S5: Abundance normalized sensitivities for selected elements to demonstrate the enhancement observed with aerosol desolvation. Sensitivities are given in cps μg^{-1} kg for two Ca matrix mass fractions to see the decrease in enhancement with higher mass load.

Elements	1 mg kg ⁻¹ Ca		50 mg kg ⁻¹ Ca	
	SN	Des	SN	Des
Li ⁺	830	10'100	630	4'000
Co ⁺	8'150	80'000	6'730	43'700
As ⁺	1'370	14'500	1'150	9'100
Pb ⁺	21'300	160'000	17'600	92'000
U ⁺	25'500	234'000	22'300	137'000

Table S6: Limits of detection and limits of quantification for elements measured with solution nebulization and desolvation. The LODs and LOQs given in ng kg⁻¹ were determined from the slope of the external calibration for the quantification study and the respective blank measurement. The dilution factor has not been considered for the LOQs.

Isotope	LODs		LOQs	
	SN	Des	SN	Des
⁷ Li ⁺	150	17	500	60
⁹ Be ⁺	26	5	90	16
¹¹ B ⁺	80	4'000	250	12'000
²³ Na ⁺	240	400	800	1'400
²⁴ Mg ⁺	15	9	50	29
²⁷ Al ⁺	24	40	80	140
³⁹ K ⁺	9	180	30	600
⁴⁰ Ca ⁺	100	250	300	900
⁵¹ V ⁺	1.5	0.6	5	1.9
⁵² Cr ⁺	2.3	0.7	8	2.6
⁵⁵ Mn ⁺	1.0	0.3	3	1.1
⁵⁶ Fe ⁺	40	15	130	50
⁵⁹ Co ⁺	4	2.0	14	7
⁶⁰ Ni ⁺	20	3	70	11
⁶³ Cu ⁺	12	6	40	21
⁶⁶ Zn ⁺	30	23	100	80
⁷⁵ As ⁺	5	6	16	20
⁸⁰ Se ⁺	11	70	40	230
⁸⁵ Rb ⁺	1.9	0.8	6	2.5
⁸⁸ Sr ⁺	1.1	0.6	4	2.0
⁹⁸ Mo ⁺	2.6	9	9	29
¹⁰⁷ Ag ⁺	2.7	2.7	9	9
¹¹¹ Cd ⁺	13	3	40	11
¹²¹ Sb ⁺	1.6	1.0	5	3
¹²⁸ Te ⁺	13	2.3	40	8
¹³⁷ Ba ⁺	6.6	1.8	22	6
²⁰⁵ Tl ⁺	1.2	0.9	4	2.9
²⁰⁸ Pb ⁺	0.8	1.9	2.6	6
²⁰⁹ Bi ⁺	0.7	1.0	2.4	3
²³⁸ U ⁺	0.3	0.1	1.1	0.4

Table S7: Quantification overview for NIST 1643f with twofold dilution for external calibration and standard addition using solution nebulization (SN) and desolvated sample introduction (Des). All values are reported in $\mu\text{g kg}^{-1}$.

Quantified Isotope	Reference		SN Ext. Cal.		SN Std. Add.		Des Ext. Cal		Des Std. Add.	
	mean	2SD	mean	2SD	mean	2SD	mean	2SD	mean	2SD
⁷ Li	16.4	0.4	18.0	0.6	16.3	0.8	15.8	0.5	17.3	0.6
⁹ Be	13.53	0.11	13.7	0.8	14.6	1.0	12.0	0.4	13.19	0.22
¹¹ B	151	7	162	5	157	5	(284)		(78)	
²³ Na	18640	240	20700	400	_b		17700	500	_b	
²⁴ Mg	7380	60	8190	150	_b		7050	240	_b	
²⁵ Mg	7380	60	7990	150	_b		7070	190	_b	
²⁷ Al	132.5	1.2	141	3	146	4	126.6	2.6	130	4
³⁹ K	1913	9	2010	50	_b		1820	50	_b	
⁴⁰ Ca	29140	300	30100	600	_b		_c		_b	
⁴¹ K	1913	9	1919	18	_b		1790	60	_b	
⁵¹ V	35.71	0.27	35.7	0.6	38.9	0.6	34.2	0.4	37.8	0.3
⁵² Cr	18.3	0.1	18.4	0.3	17.9	0.6	18.0	0.4	19.40	0.20
⁵³ Cr	18.3	0.1	18.0	0.6	18.38	0.22	17.8	0.5	19.2	0.4
⁵⁵ Mn	36.8	0.6	36.0	0.6	36.7	0.8	35.0	1.2	34.0	0.6
⁵⁶ Fe	92.5	0.8	100	3	95.9	1.8	90.1	3.0	87.5	1.2
⁵⁷ Fe	92.5	0.8	185	5	266	5	147	10	155.6	2
⁵⁹ Co	25.05	0.17	25.0	0.5	25.6	0.4	23.7	0.8	25.1	0.3
⁶⁰ Ni	59.2	1.4	58.8	2.0	61.4	1.0	55.9	1.8	66.6	1.2
⁶² Ni	59.2	1.4	56.7	2.2	60.0	1.4	56.4	0.4	65.1	1.0
⁶³ Cu	21.4	0.7	19.1	0.4	22.4	0.8	19.8	0.5	22.5	0.5
⁶⁴ Zn	73.7	1.7	85.4	2.8	79.4	1.2	77.6	2.4	86.6	1.4
⁶⁵ Cu	21.4	0.7	19.3	0.4	22.5	0.5	19.5	0.6	22.7	0.4
⁶⁶ Zn	73.7	1.7	85	3	75.6	1.6	76.4	2.6	83.0	1.4
⁷⁵ As	56.9	0.4	56.3	1.2	59.8	0.8	53.8	1.0	54.8	0.5
⁷⁸ Se	11.58	0.08	11.2	0.8	11.6	0.8	11.43	0.24	12.4	0.4
⁸⁰ Se	11.58	0.08	11.27	0.26	11.3	0.6	10.69	0.24	12.0	0.4
⁸⁵ Rb	12.51	0.12	12.33	0.20	13.2	0.3	12.00	0.30	13.3	0.3
⁸⁸ Sr	311	18	329	4	317	5	298	6	288.4	2.4
⁹⁵ Mo	114.2	1.7	110.8	1.8	114.4	2.2	109	3	114.4	1.8
⁹⁸ Mo	114.2	1.7	111.1	2.0	114.8	2.4	110	4	114.0	2.2
¹⁰⁷ Ag	0.961	0.008	0.96	0.04	0.94	0.04	0.966	0.022	0.905	0.026
¹⁰⁹ Ag	0.961	0.008	1.03	0.03	1.00	0.06	1.030	0.028	0.982	0.020
¹¹¹ Cd	5.8	0.13	6.12	0.28	5.77	0.24	5.84	0.22	6.09	0.14
¹²¹ Sb	54.9	0.39	50.0	1.2	54.7	1.0	49.1	1.6	59.3	0.6
¹²⁸ Te	1.0	0.0082	0.89	0.03	0.95	0.05	0.889	0.012	1.046	0.024
¹³⁷ Ba	513	7	510	6	535	6	498	16	512	8
¹³⁸ Ba	513	7	540	14	500	8	490	14	511	16
²⁰⁵ Tl	6.82	0.03	6.84	0.22	7.32	0.26	7.0	0.4	7.32	0.18
²⁰⁸ Pb	18.30	0.08	17.67	0.20	18.9	0.5	17.7	0.8	18.13	0.22
²⁰⁹ Pb	12.50	0.10	12.6	0.3	13.3	0.5	12.3	0.4	13.21	0.18
²³⁸ U	_a		0.0063	0.0013	0.0052	0.0010	0.0055	0.0004	0.0049	0.0006

^a not available, ^b not determined, ^c detector saturation

Table S8: Quantification overview for NIST 1643f with tenfold dilution for external calibration and standard addition using solution nebulization (SN) and desolvated sample introduction (Des). All values are reported in $\mu\text{g kg}^{-1}$.

Quantified Isotope	Reference		SN Ext. Cal.		SN Std. Add.		Des Ext. Cal		Des Std. Add.	
	mean	2SD	mean	2SD	mean	2SD	mean	2SD	mean	2SD
⁷ Li	16.4	0.4	15.8	0.8	16.7	0.8	17.4	1.2	15.7	0.6
⁹ Be	13.53	0.11	12.3	2.0	13.0	1.8	12.4	0.8	12.4	0.4
¹¹ B	151	7	150	10	141	5	(128)		(124)	
²³ Na	18640	240	19900	1100	- ^c		18900	1400	- ^c	
²⁴ Mg	7380	60	6910	180	- ^c		7390	260	- ^c	
²⁵ Mg	7380	60	7740	240	- ^c		7480	200	- ^c	
²⁷ Al	132.5	1.2	134	5	144.6	3.0	130	4	124	2.2
³⁹ K	1913	9	2040	60	- ^c		1890	70	- ^c	
⁴⁰ Ca	29140	300	29700	800	- ^c		29200	1000	- ^c	
⁴¹ K	1913	9	1950	50	- ^c		1910	100	- ^c	
⁵¹ V	35.71	0.27	34.5	1.2	36.27	0.26	36.2	1.0	35.31	0.26
⁵² Cr	18.3	0.1	18.3	0.8	18.7	0.3	18.7	0.4	18.06	0.12
⁵³ Cr	18.3	0.1	17.9	0.6	18.3	1.2	18.1	0.3	17.62	0.22
⁵⁵ Mn	36.8	0.6	35.9	1.0	36.23	0.16	35.8	0.8	34.9	0.4
⁵⁶ Fe	92.5	0.8	95.4	2.6	96.6	1.0	92	4	83.5	2.4
⁵⁷ Fe	92.5	0.8	172	8	257.2	6	141	12	151	3
⁵⁹ Co	25.05	0.17	24.4	0.8	25.3	0.4	25.5	1.2	27.0	0.5
⁶⁰ Ni	59.2	1.4	56.8	1.8	60.5	2.0	59.2	2.2	63.1	1.4
⁶² Ni	59.2	1.4	57.1	2.4	58.5	1.4	59.0	2.6	62.4	1.8
⁶³ Cu	21.4	0.7	18.6	1.0	21.5	0.4	19.9	0.4	23.17	0.28
⁶⁴ Zn	73.7	1.7	75.8	1.8	75.0	1.4	76.2	2.2	83.2	1.8
⁶⁵ Cu	21.4	0.7	19.0	0.6	22.0	1.0	19.6	0.8	22.2	0.4
⁶⁶ Zn	73.7	1.7	75.6	2.6	72.1	2.0	74.5	3	78.3	2.0
⁷⁵ As	56.9	0.4	55.0	2.2	55.5	1.8	52.4	1.2	55.9	0.5
⁷⁸ Se	11.58	0.08	11.0	1.8	12.3	3.0	12.1	0.8	10.5	0.4
⁸⁰ Se	11.58	0.08	11.1	1.2	11.3	1.4	12.1	0.5	10.1	0.8
⁸⁵ Rb	12.51	0.12	12.7	0.5	12.80	0.18	12.62	0.14	13.39	0.16
⁸⁸ Sr	311	18	291	8	320	6	291	6	293	10
⁹⁵ Mo	114.2	1.7	107.4	2.2	112.0	1.6	109.3	2.0	106.3	1.8
⁹⁸ Mo	114.2	1.7	107.3	3.0	114.0	2.2	109	5	101.9	2.0
¹⁰⁷ Ag	0.961	0.008	0.97	0.06	0.95	0.10	0.95	0.03	0.88	0.03
¹⁰⁹ Ag	0.961	0.008	1.02	0.06	1.00	0.10	1.01	0.03	0.935	0.024
¹¹¹ Cd	5.8	0.13	5.9	0.4	6.0	0.5	5.9	0.3	5.9	0.3
¹²¹ Sb	54.9	0.39	48.3	2.0	54.2	0.4	49.1	1.2	59.1	0.8
¹²⁸ Te	1.0	0.0082	0.85	0.12	0.94	0.08	0.83	0.08	1.05	0.06
¹³⁷ Ba	513	7	510	5	558	5	504	20	468	8
¹³⁸ Ba	513	7	521	16	528	4	479	26	463	16
²⁰⁵ Tl	6.82	0.03	6.80	0.22	7.02	0.22	7.4	0.3	7.27	0.22
²⁰⁸ Pb	18.30	0.08	17.66	0.22	19.2	0.4	18.6	0.8	18.1	0.6
²⁰⁹ Pb	12.50	0.10	12.6	0.4	12.8	0.4	12.4	0.8	13.3	0.6
²³⁸ U	- ^a		- ^b		- ^b		0.0052	0.0006	0.0067	0.0014

^a not available, ^b below LOQ, ^c not determined




Figure S3: Quantification overview obtained with a twofold dilution of SLRS-6 for solution nebulization (black) and aerosol desolvation (red) using external calibration. For each isotope, the deviation of the obtained mass fractions from the certified reference material is given with error bars representing two times the standard deviation. The grey horizontal lines correspond to a difference of $\pm 10\%$ from the certified value.




Figure S4: Quantification overview obtained with a twofold dilution of SLRS-6 for solution nebulization (black) and aerosol desolvation (red) using standard addition. For each isotope, the deviation of the obtained mass fractions from the certified reference material is given with error bars representing two times the standard deviation. The grey horizontal lines correspond to a difference of $\pm 10\%$ from the certified value.




Figure S5: Quantification overview obtained with a tenfold dilution of NIST-1643f for solution nebulization (black) and aerosol desolvation (red) using external calibration. For each isotope, the deviation of the obtained mass fractions from the certified reference material is given with error bars representing two times the standard deviation. The grey horizontal lines correspond to a difference of $\pm 10\%$ from the certified value.




Figure S6: Quantification overview obtained with a tenfold dilution of NIST-1643f for solution nebulization (black) and aerosol desolvation (red) using standard addition. For each isotope, the deviation of the obtained mass fractions from the certified reference material is given with error bars representing two times the standard deviation. The grey horizontal lines correspond to a difference of $\pm 10\%$ from the certified value.