



Fig. G Calibration line of signal intensity for ⁹⁸Mo using the selected reference materials. The error bars represent the standard deviation of 6 measurements. Using this calibration line, Mo in NIST 1568a was calculated to be 1.44 mg kg⁻¹, which was 99% of the certified value.